

**WAC 296-307-63630 Make sure a record is kept of audiometric tests.**

**You must:**

• Retain a legible copy of all employee audiograms conducted under this part.

- Make sure the record includes:

■ Name and job classification of the employee

■ Date of the audiogram

■ The examiner's name

■ Date of the last acoustic or exhaustive calibration of the audiometer

■ Employee's most recent noise exposure assessment

■ The background sound pressure levels in audiometric test rooms.

[Statutory Authority: RCW 49.17.010, 49.17.040, 49.17.050, 49.17.060. WSR 05-01-166, § 296-307-63630, filed 12/21/04, effective 4/2/05.]