

WAC 246-240-632 Records of five-year inspection for teletherapy and gamma stereotactic radiosurgery units. (1) A licensee shall maintain a record of the five-year inspections for teletherapy and gamma stereotactic radiosurgery units required by WAC 246-240-393 for the duration of use of the unit.

(2) The record must contain:

(a) The inspector's radioactive materials license number;

(b) The date of inspection;

(c) The manufacturer's name and model number and serial number of both the treatment unit and source;

(d) A list of components inspected and serviced, and the type of service; and

(e) The signature of the inspector.

[Statutory Authority: RCW 70.98.050. WSR 06-05-019, § 246-240-632, filed 2/6/06, effective 3/9/06.]