

**WAC 246-243-070 Storage precautions.** (1) Locked radiographic exposure devices and storage containers shall be physically secured to prevent tampering or removal by unauthorized personnel.

(2) At least one calibrated and operable radiation survey instrument shall be available at the storage area whenever a radiographic exposure device, a storage container, or source is being placed in storage.

[Statutory Authority: RCW 70.98.050. WSR 94-01-073, § 246-243-070, filed 12/9/93, effective 1/9/94. Statutory Authority: RCW 43.70.040. WSR 91-02-049 (Order 121), recodified as § 246-243-070, filed 12/27/90, effective 1/31/91. Statutory Authority: RCW 70.98.050. WSR 81-01-011 (Order 1570), § 402-36-050, filed 12/8/80; Order 1084, § 402-36-050, filed 1/14/76; Order 1, § 402-36-050, filed 1/8/69; Rules (part), filed 10/26/66.]