

**Chapter 246-225A WAC**  
**RADIATION SAFETY AND DIAGNOSTIC IMAGE QUALITY STANDARDS FOR DENTAL FACILITIES**

Last Update: 9/7/11

**WAC**

246-225A-001	Purpose and scope.
246-225A-010	Definitions.
246-225A-020	General requirements and administrative controls.
246-225A-025	X-ray system radiation safety procedure.
246-225A-026	Healing arts screening program.
246-225A-040	Dental X-ray rule variance request.
246-225A-050	Dental X-ray facility design.
246-225A-060	General requirements for all dental X-ray systems.
246-225A-070	Special requirements for dental extra-oral radiography.
246-225A-080	Special requirements for dental intra-oral radiography.
246-225A-085	Hand-held X-ray system.
246-225A-090	X-ray image processing requirements.
246-225A-110	Film processing quality assurance.

**WAC 246-225A-001 Purpose and scope.** This chapter establishes facility design and operation requirements for the use of dental X-ray equipment according to chapter 70.98 RCW. The scope of this chapter pertains to dental intra-oral and extra-oral radiography and establishes radiation safety requirements for patients, dental employees, and the public; and establishes optimal diagnostic image processing requirements.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 08-14-074, § 246-225A-001, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-010 Definitions.** The definitions in this section apply throughout this chapter unless the context clearly requires otherwise.

(1) "Artifact" means an undesirable optical density or blemish on a radiographic image that detracts from the diagnostic information.

(2) "Barrier" (see "protective barrier").

(3) "Beam" (see "X-ray").

(4) "Beam-limiting device," sometimes called a collimator or cone, means a device that controls the size of the X-ray field.

(5) "Cephalometric" means X-ray imaging specific to the human head and jaw.

(6) "Control panel" means the part of the X-ray system where the switches, knobs, pushbuttons, and other hardware necessary to operate the X-ray system are located.

(7) "CR (computed radiography)" means creating an X-ray image using plates consisting of a special phosphor that when exposed to radiation and then processed by a scanner, provides the information to a computer for display and manipulation.

(8) "CT (computed tomography)" means creating a cross-sectional X-ray image generated by an X-ray source and detector moving around the patient's body.

(9) "Dead-man button" means an X-ray exposure button designed so that it can only be operated by continuous pressure on the button by the operator, and when released before the preset exposure time will stop the exposure.

(10) "Department" means the department of health, which is the state radiation control agency under chapter 70.98 RCW.

(11) "Detector" means a device capable of receiving and recording an X-ray image.

- (12) "Diagnostic source assembly" means the combination of the tube housing assembly and the collimator.
- (13) "Direct scattered radiation" means radiation discharged in a straight line from the object being radiographed.
- (14) "DR (direct digital radiography)" means creating an X-ray image by sending signals directly from a solid state detector to a computer for display and manipulation.
- (15) "Exposure," as the context implies, means:
- (a) The number of electrons, measured in coulombs per kilogram of air, released through the ionization of air molecules by electromagnetic radiation; or
- (b) An occupational worker or patient being subjected to radiation either directly or indirectly.
- (16) "Extra-oral radiography" means creating a film or digital X-ray image on an image receptor placed outside the mouth. Examples include panoramic and cephalometric X-rays.
- (17) "Filter" means material, such as copper or aluminum, placed in the useful beam of the X-ray to block selected energies, and in a safelight to block light that could fog the X-ray film.
- (18) "Floor plan" means a drawing of the X-ray room, along with its dimensions, identification of adjacent areas and occupiable space above and below.
- (19) "Focal spot" means the area on the anode end of the X-ray tube bombarded by the electrons accelerated from the cathode and from which the useful X-ray beam begins.
- (20) "Grid" means a device placed between the patient and the image receptor in extra-oral radiography that reduces scattered radiation that would decrease the quality of the image being created.
- (21) "Hand-held" (see "X-ray system").
- (22) "Healing arts screening" means using X-ray equipment without an order by a licensed practitioner on an individual who does not have a known or diagnosed disease or symptom to learn if the individual may have an indication of ill health.
- (23) "HVL (half-value layer)" means the thickness of material that reduces the intensity of radiation to one-half of its original value.
- (24) "Image receptor" means a device that captures an X-ray beam for image processing.
- (25) "Intra-oral radiography" means creating a film or digital X-ray image on an image receptor placed inside the mouth.
- (26) "kV (kilovolt)" means the unit used to measure electrical energy.
- (27) "kVp (kilovolts peak)" means the highest possible voltage across the X-ray tube during an exposure (see also "peak tube potential").
- (28) "Leakage radiation" means radiation coming from the X-ray tube, other than the useful X-ray beam.
- (29) "Leakage technique factors" means the technique factors associated with the tube housing assembly that are used to measure leakage radiation. They are defined as the maximum rated peak tube potential and the maximum rated continuous tube current at the maximum peak tube potential.
- (30) "Licensed practitioner" means an individual licensed to practice dentistry under chapter 18.32 RCW.
- (31) "mA (milliampere)" means the unit used to measure electrical current in an X-ray tube.

(32) "mAs (milliampere second)" means the product of the electrical current in the X-ray tube in mA and the time of exposure in seconds.

(33) "Mobile" (see "X-ray system").

(34) "Operator" means a person working under the direction of a licensed practitioner to operate X-ray equipment and who has been properly trained according to WAC 246-225A-020.

(35) "Operatory" means a room in which dental health care procedures are performed.

(36) "Peak tube potential" means the maximum voltage in the X-ray tube during an exposure.

(37) "Portable" (see "X-ray system").

(38) "Position-indicating device" means a device on X-ray equipment that shows where the X-ray beam will be directed and establishes the distance from the X-ray tube to the patient's body. The device may or may not incorporate or serve as a beam-limiting device.

(39) "Primary beam" (see "useful beam").

(40) "Primary protective barrier" means the material placed in the useful beam, beyond the patient and image receptor, to reduce remnant primary beam exposure.

(41) "Protected area" means a space for X-ray equipment operators that is shielded so that X-ray exposures are reduced enough to meet the exposure limits of WAC 246-221-010 (Occupational dose limits for adults) and WAC 246-220-007 (Statement of philosophy). In addition, the space must have no exposure to direct scattered radiation.

(42) "Protective apron" means a garment made of radiation absorbing materials used to reduce a person's radiation exposure.

(43) "Protective barrier" means a structure made of radiation absorbing material used to reduce radiation exposure.

(44) "Quality assurance" means a program designed to produce high quality X-ray images at minimal cost and with minimal patient exposure to radiation.

(45) "Quick developer" means small-volume chemistry designed to process dental intra-oral film in less than a minute.

(46) "Radiation safety" means ways to protect patients and staff from unnecessary radiation exposure. Safety measures may include patient exposure reduction, image quality improvement, diagnostic imaging system quality assurance, radiation measurements, dose evaluations, compliance with state and federal regulations, and related issues.

(47) "Radiographic" means the production of an image created when an X-ray pattern exits an X-rayed object.

(48) "Radiography" means a way of creating a permanent film or digital image using X-rays.

(49) "Recording" means creating a permanent image, on film or in a computer, from an X-ray exposure.

(50) "Registrant" means the owner or controller of the radiation equipment who is responsible for the safe operation of the radiation equipment in accordance with this chapter and chapter 70.98 RCW.

(51) "Registration" means providing required information and continuing contact with the department by any person possessing a radiation machine in accordance with chapter 246-224 WAC, Radiation protection—Radiation machine assembly and registration.

(52) "Remnant primary beam" means the part of the useful beam that completely passes through the patient and image receptor.

(53) "Ring-detector type CT" means computed tomography performed with a fan-shaped beam that generates image slices of anatomy rather than using a cone-shaped beam creating a volumetric picture.

(54) "Safelight" means a lamp with a filter that is used in an X-ray darkroom to provide enough light to see, but not enough to fog the film.

(55) "Scattered radiation" means radiation that has changed direction, or generated other radiation as it impacts or passes through matter.

(56) "Scram button" means a large, prominently displayed button, mounted in an X-ray operator's area to allow quick termination of an X-ray exposure in case of an emergency.

(57) "Secondary protective barrier" means an object or material sufficient to reduce stray radiation to the required degree as stated in chapter 246-221 WAC (Radiation protection standards).

(58) "SID (source-to-image-receptor distance)" means the distance from the focal spot in the X-ray tube to the center of the surface of the image receptor.

(59) "Source" means the focal spot of the X-ray tube.

(60) "SSD (source-to-skin distance)" means the distance between the focal spot of the X-ray tube and the nearest point on the patient's skin where the primary beam enters.

(61) "Stationary" (see "X-ray system").

(62) "Stray radiation" means the sum of leakage and scattered radiation.

(63) "Technique chart" means a written instruction or guide that X-ray equipment operators use to determine which radiation technique factors to select for each type of radiographic examination.

(64) "Technique factors" means the X-ray system settings selected for a given radiographic examination. They are specified as the peak tube potential in kVp and either:

(a) Tube current measured in mA and exposure time in seconds or pulses; or

(b) The product of tube current and exposure time expressed in mAs.

(65) "Tube" means a glass tube that produces an X-ray when high-voltage electricity is passed between the cathode at one end and the anode at the other.

(66) "Tube housing assembly" means the X-ray tube and its housing. It includes high-voltage or filament transformers and other appropriate elements when they are contained within the tube housing.

(67) "Tube housing port" means the portion of the tube housing assembly that the X-rays pass through.

(68) "Useful beam" means the radiation that passes through the tube housing port and the opening of the beam-limiting device.

(69) "Variance" means a department-authorized alternative to a requirement of this chapter.

(70) "X-ray" means a beam of ionizing radiation produced by a machine.

(71) "X-ray control" means a device that controls how much electricity enters the X-ray high-voltage generator or the X-ray tube. It includes equipment that controls the technique factors for an exposure.

(72) "X-ray equipment" means the entire X-ray system or parts of the system.

(73) "X-ray exposure button" means the part of the X-ray system that when engaged generates the production of an X-ray.

(74) "X-ray high-voltage generator" means a device that supplies electrical energy to the X-ray tube to create an X-ray beam.

(75) "X-ray system" means all of the components of a machine used for the controlled production of X-rays. It includes minimally an X-ray high-voltage generator, an X-ray control, a tube housing assembly, a beam-limiting device, and the necessary supporting structures. Additional components which function with the system, such as the image receptor, are considered integral parts of the system. Types of X-ray systems are:

(a) "Hand-held" means a self-contained X-ray system designed to be held in one or two hands to perform intra-oral radiography. Hand-held X-ray systems used on a tripod or stand are considered to be "portable" systems.

(b) "Mobile" means an X-ray system mounted on a permanent base with wheels or casters for moving the X-ray system fully assembled. It is intended to be taken from one geographical location to another or from one room to another.

(c) "Portable" means an X-ray system designed to be hand-carried, but not hand-held during use.

(d) "Stationary" means an X-ray system that is installed in a fixed location, such as bolted to a floor or wall.

(76) "X-ray tube" means any electron tube which is designed to be used primarily for the production of X rays.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-010, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-010, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-020 General requirements and administrative controls.** The registrant is responsible for directing the operation of the X-ray system and assuring the provisions of chapter 246-222 WAC (Radiation protection—Worker rights) are met. In addition, the registrant shall:

(1) Verify that any operator of the X-ray equipment is trained and able to show that he or she can correctly and safely operate the X-ray equipment used by the registrant. The department may determine compliance by observation, interview, or testing in these subject areas:

(a) Knowledge of the X-ray system controls and their function;

(b) Knowledge of radiation safety and shielding methods for both operators and patients;

(c) Proper image processing.

(2) Post a technique chart at each X-ray system's control panel that specifies the following information for the examinations being performed by that system:

(a) Patient's teeth, jaw, or head anatomy versus technique factors to be used;

(b) If applicable, settings for automatic exposure devices; and

(c) The type and size of screen-film combination or other imaging system to be used.

(3) Require that all individuals, other than the patient being examined:

(a) Be positioned so that no part of the body, including the extremities, will be struck by the useful beam;

(b) Be protected from stray radiation by wearing protective aprons or by being positioned behind protective barriers of not less than 0.25 millimeters lead equivalent; and

(c) Not be present in the room during the X-ray exposure, except:

(i) As described in subsection (4)(b) of this section; or

(ii) When a hand-held, portable, or mobile X-ray system is used.

(4) Use mechanical holding devices when a patient, film, or image receptor needs to be supported during an X-ray exposure when the technique permits.

(a) An individual may not be allowed to routinely hold a patient, film, or image receptor; and

(b) Holding a patient, film, or image receptor must only be allowed in very unusual and rare situations. In these cases the patient's name, the date, and the name of the person holding the patient must be recorded in writing and maintained by the registrant for at least five years.

(5) Comply with the occupational exposure limits and the requirements for the determination of prior occupational dose stated under WAC 246-221-020 (Determination of prior occupational dose) for all individuals associated with the operation of the registrant's X-ray system. In addition, when protective clothing or devices are worn on portions of the body and a dosimeter is required, at least one dosimeter must be used and documented as follows:

(a) When an apron is worn, the dosimeter must be worn at the collar outside the apron;

(b) The dose to the whole body based on the maximum dose attributed to the most critical organ must be recorded on the reports required under WAC 246-221-230 (Records important to radiation safety). If more than one dosimeter is worn, each dose must be identified with the area where the dosimeter was worn on the body.

(6) Require personnel dosimetry of an operator when:

(a) Mobile, portable, or hand-held X-ray systems are used, i.e., when X-ray exposure buttons or X-ray exposure button cords are used that allow the operator to stand in an unprotected area during exposures; and

(b) Measurements by the department show ten percent of the exposure limits as specified under WAC 246-221-010 (Occupational dose limits for adults) are exceeded.

(7) Use only X-ray equipment, and the accessories used in connection with making X-rays, that meet the requirements of this chapter.

(8) Not allow anyone in the dental office to operate X-ray equipment for diagnostic purposes when the X-ray equipment:

(a) Does not meet the provisions of this chapter; or

(b) Is malfunctioning or threatens the health or safety of a patient, dental employee, or the public.

(9) Not allow patients to be exposed to the useful X-ray beam except for healing arts purposes. Only a licensed practitioner may authorize an exposure to the useful beam. Deliberate exposure of an individual for the following purposes is prohibited:

(a) Training, demonstration, or other purposes unless there are also healing arts requirements and proper prescription provided; or

(b) Except for exposure required under medicare provisions, any exposure for which the sole purpose is satisfying a third party's prerequisite for reimbursement under any health care plan.

(10) Submit shielding specifications designed by a qualified expert as defined in chapter 246-220 WAC and floor plans to the depart-

ment for review if the registrant proposes to use ring-detector type CT or medical X-ray systems for dental imaging. The submittal must:

(a) Meet the requirements of WAC 246-225A-050; and

(b) Be based on the criteria and methods found in National Council on Radiation Protection and Measurements (NCRP) report #147, *Structural Shielding Design for Medical X-Ray Imaging Facilities*, issued November 19, 2004.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-020, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-020, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-025 X-ray system radiation safety procedure. (1)**

The department may require the registrant to adopt a written X-ray system radiation safety procedure if there is reason to believe the registrant needs increased attention because of:

(a) Inadequate operator training;

(b) Extremely high workload;

(c) Increased risk of exposure due to staff supporting patients during radiography;

(d) Increased risk of exposure to scattered radiation;

(e) Unnecessarily high patient exposure values; or

(f) Other similar conditions.

(2) The X-ray system radiation safety procedures must:

(a) Address patient and occupationally exposed personnel safety; and

(b) Define any restrictions of the operating technique required for safe operation of the X-ray system.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-025, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-025, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-026 Healing arts screening program.** Any individual proposing to conduct a healing arts screening program shall obtain approval from the state health officer as required in WAC 246-225-99930 before conducting the screening program.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 08-14-074, § 246-225A-026, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-040 Dental X-ray rule variance request.** A registrant may submit a written request to the department for a variance from the applicable regulations. The registrant shall not use X-ray equipment on patients until the department approves the variance request.

(1) The written request shall be addressed to: X-ray Supervisor, Office of Radiation Protection, Department of Health, P.O. Box 47827, Olympia, Washington 98504-7827, and must include:

(a) The specific WAC reference or references of the rule for which the variance is requested;

(b) An explanation of the circumstances involved, and the reason why the rule cannot be followed;

(c) A description of how the proposed alternative meets the intent of the rule and how the registrant shall protect patients, dental employees, and the public;

(d) A description of the X-ray system to be used with supporting pictures or documents; and

(e) The time period for which the variance is requested.

(2) The department may impose conditions that may be necessary to protect human health and safety during the term of the variance.

(3) If necessary, the department may require the registrant to submit additional information.

(4) The department may conduct an on-site variance inspection to verify the information provided or if it determines that an inspection is necessary.

(5) As determined by the department, variances can be permanent or temporary.

(6) The department may at any time revoke a variance if it is determined that the conditions of the variance are not being followed.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 08-14-074, § 246-225A-040, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-050 Dental X-ray facility design.** (1) Each X-ray exposure button must be located to meet the following criteria:

(a) For stationary X-ray systems, the X-ray exposure button must be permanently mounted in a protected area so that the operator can make an exposure only from the protected area; and

(b) Mobile X-ray systems must have an X-ray exposure button located at the end of a cord at least twelve feet (3.7 meters) long.

(2) Shielding for cephalometric X-ray must meet the following criteria:

(a) Be at least one foot (30.5 centimeters) larger, in both the horizontal and vertical directions, than the area of the primary beam where it strikes the nearest wall; and

(b) Shielding between the nearest wall struck by the primary beam and the next occupied area must have two-pound lead or equivalent installed in the wall. Exterior walls or concrete block walls need no additional shielding.

(3) Acceptable shielding materials for dental X-ray facilities are as follows:

(a) The minimum shielding for intra-oral stray radiation protection is standard gypsum wallboard/sheetrock construction (two layers each of five-eighths inch thickness).

(b) Where windows are provided to observe patients during radiography, the windows are at least one-half inch plate glass, or equivalent ability to reduce exposure.

(c) All other materials used for shielding between operatories and for operator protection areas are equivalent to 0.2 millimeters of lead.

(4) Barriers surrounding dental X-ray rooms and dental operatories where intra-oral X-ray equipment is installed must meet the following criteria:

(a) Be at least six feet (1.83 meters) high and composed of materials capable of reducing scattered radiation as required under subsection (3) of this section;



(b) There must be no line of sight between workers or patients in one operatory and the X-ray tube housing assembly in the next operatory when that X-ray tube housing assembly is in its operating position;

(c) X-ray tube housing assemblies must not be mounted between operatories on top of barriers less than six feet (1.83 meters) high, unless those barriers are at the foot end of the patient chairs, and there is no line of sight between adjacent operatories.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-050, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-050, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-060 General requirements for all dental X-ray systems.** Registrants shall use only dental X-ray systems and medical X-ray systems for dental imaging that meet the following requirements:

(1) The leakage radiation from the tube housing assembly, measured at a distance of one meter in any direction from the source, must not exceed 100 milliroentgens in one hour when the X-ray tube is operated at its leakage technique factors. The department will determine compliance by measuring leakage averaged over an area of 100 square centimeters with no dimension of that area greater than 20 centimeters.

(2) The HVL of the useful beam for a given X-ray tube potential must not be less than the values shown in Table 1 of this section. To determine the HVL at an X-ray tube potential which is not listed in Table 1 of this section, linear interpolation or extrapolation may be made.

Table 1

Design operating range (kVp)	Measured potential (kVp)	HVL (millimeters of aluminum equivalent)
70 and below	70 and below	1.5
Above 70	71	2.1
	80	2.3
	90	2.5
	100	2.7

(3) If two or more X-ray tubes are controlled by one X-ray exposure button, the tube or tubes in operation must be clearly marked before an exposure, on both the X-ray control panel and near or on the selected tube housing assembly.

(4) The tube housing assembly supports of a stationary, portable, or mobile X-ray system must be adjusted so that the tube housing assembly remains stable and does not drift during an exposure unless the tube housing movement during exposure is a designed function of the X-ray system. Except for X-ray systems specifically designed to be hand-held, an X-ray system or tube housing assembly must not be hand-held by anyone during the exposure.

(5) Except for CT X-ray systems that have a scram button, each X-ray control must have a dead-man button.

(6) Technique indicators must be set as follows:

(a) All exposure technique factors must be set on the control panel before the exposure begins, except when automatic exposure con-

trols are used. When automatic exposure controls are used, any preselected settings for each exposure must be indicated.

(b) On equipment having fixed technique factors, the requirement in (a) of this subsection may be met by permanent markings or labels.

(7) Linearity must be measured and met as follows:

(a) The difference between the ratio of milliroentgens (mR) exposure to mAs at one mA or mAs setting and the ratio of mR exposure to mAs at another mA or mAs setting must not exceed 0.1 times the sum of the ratios. This is written as:

$$X_1 - X_2 \leq 0.10 (X_1 + X_2)$$

Where  $X_1$  and  $X_2$  are the ratios (mR/mAs) for each mA or mAs setting.

(b) The measurement must be performed at any selection of mA or mAs without regard to focal spot size, provided neither focal spot size is less than 0.45 millimeters.

(8) When four exposures are made at identical operating settings, the difference between the maximum exposure ( $E_{max}$ ) and the minimum exposure ( $E_{min}$ ) must be less than or equal to ten percent of the average exposure ( $E$ ). This is written as:

$$(E_{max} - E_{min}) \leq 0.1E$$

(9) The difference between the kVp indicated on an X-ray system and the measured kVp must not be greater than ten percent of the indicated kVp.

(10) Timers must be able to:

(a) Stop the exposure at a preset time interval, a preset product of current and time, a preset number of pulses, or a preset radiation exposure to the image receptor; and

(b) Reset automatically to the initial setting or to zero when the exposure is stopped.

(11) X-ray equipment must not be capable of making an exposure when the timer is set to the zero or off position if either position is provided.

(12) Each X-ray control must have a visual indicator (such as a light) or audible signal so that the operator knows that X-rays are being produced or the exposure is occurring or has ended.

(13) Registrants shall not use dental fluoroscopy without electronic amplification.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-060, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-060, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-070 Special requirements for dental extra-oral radiography.** Registrants shall use X-ray systems for extra-oral radiography that meet the following requirements for:

(1) Beam limitation.

(a) X-ray equipment designed for only one image receptor size at a fixed SID must be able to limit the size of the beam at the plane of the image receptor to no larger than the image receptor, and to align the center of the X-ray beam with the center of the image receptor to within two percent of the SID. In the case of extra-oral imaging systems where the image receptor can be turned vertically or horizontally, the beam-limiting device must also be able to be turned so that

the dimensions of the beam match the image receptor dimensions at the image receptor plane.

(b) Intra-oral radiography systems used to perform cephalometric projections, including trans-cranial exams, must be equipped with a stable means to:

(i) Comply with the beam size dimensions in subsection (1)(a) of this section; and

(ii) Center the beam to the image receptor as required in subsection (1)(a) of this section.

(c) General purpose medical X-ray equipment used to perform dental radiography must:

(i) Have stepless adjustment of the dimensions of the X-ray beam so that the width and height of the X-ray beam are independently adjustable. The minimum beam size at an SID of 100 centimeters must be equal to or less than 10 by 10 centimeters.

(ii) Have a means for operators to visually set the width and height of the X-ray beam. The misalignment of the edges of the visually set light field with the respective edges of the X-ray beam along either the length or width of the visually set light field must not be more than two percent of the distance from the source to the center of the visually defined light field when the surface upon which it appears is perpendicular (at a 90 degree angle) to the central axis of the X-ray beam.

(iii) Have a means to indicate on the X-ray equipment when the axis of the X-ray beam is perpendicular to the plane of the image receptor and to align the center of the X-ray beam to the center of the image receptor to within two percent of the SID (five percent for equipment manufactured before August 1974). Dental lateral jaw examinations are excluded from this requirement.

(iv) Have a beam-limiting device that indicates the X-ray beam size in centimeters or inches at the plane of the image receptor to which the beam-limiting device is adjusted.

(v) Have an actual beam size at the plane of the image receptor that matches the indicated size to within two percent of the SID.

(2) SSD and SID.

(a) Dental extra-oral radiography systems must have a durable, securely fastened means to limit the SSD to not less than 23 centimeters. The requirement may be met when the beam-limiting device provides the required limits.

(b) Dental extra-oral radiography systems in which the SID is not fixed (such as an intra-oral system used for cephalometrics) must have a device or reference that will indicate the actual SID to within two percent of the indicated SID.

(3) Viewing device.

Dental extra-oral radiography installations must provide a viewing device (mirror, video camera, or glass window designed to reduce exposure) so that operators of the X-ray equipment may observe the patient's head and neck area during the exposure without being exposed to the primary beam or stray radiation.

(4) Scattered radiation suppressing grids. When using scattered radiation suppressing grids, the grids must be:

(a) Clearly labeled with the SID for which the grids are designed to be used; and

(b) Used at the proper SID.

(5) X-ray film and screen requirements.

(a) X-ray film used for extra-oral imaging must be used before the expiration date specified by the manufacturer.

(b) The spectral sensitivity of the X-ray film used must be matched by the appropriate spectral output of the intensifying screens used in the cassettes as recommended by the film and screen manufacturers.

(c) Screens must be free of dirt, abrasions, and discoloration that would cause artifacts on the image.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-070, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-070, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-080 Special requirements for dental intra-oral radiography.**

(1) Registrants using an X-ray system designed for use with an intra-oral image receptor shall use equipment that:

- (a) Limits the SSD to not less than 18 centimeters;
- (b) Limits the X-ray beam so that the beam diameter at the minimum SSD is no greater than 7 centimeters in diameter;
- (c) Has an open-ended position-indicating device; and
- (d) Has shielding included in the beam-limiting device or position-indicating device equivalent to that required for the diagnostic source assembly under WAC 246-225A-060(1).

(2) Registrants shall not use diagnostic dental X-ray systems with a fixed, nominal kVp of less than 55.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-080, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-080, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-085 Hand-held X-ray system.** The following requirements apply to hand-held X-ray systems:

(1) Registrants using hand-held X-ray systems must provide for security and safe storage while not in use.

(2) The image receptor used with hand-held dental X-ray systems must either be:

- (a) A speed class of intra-oral film designated as "F," "E/F" or faster; or
- (b) A digitally acquired image (CR or DR).

(3) The hand-held X-ray system must be equipped with a backscatter shield of not less than 0.25 mm lead equivalent and 15.2 cm (6 inches) in diameter that can be positioned to within 1 cm of the end of the position indicating device. The hand-held X-ray system must always be used with the backscatter shield in place.

(4) Conditions and restrictions using the predicted whole body dose rate (effective dose) to the operator and shallow dose to the fingers:

<b>If the predicted whole body dose rate (effective dose) is:</b>	<b>And the shallow dose rate to the fingers is:</b>	<b>The operator:</b>
<100 mrem/yr	<1 rem/yr	Can operate equipment with no additional requirements
>100 mrem/yr	<1 rem/yr	Shall wear a leaded apron of 0.25 mm lead equivalent
>100 mrem/yr	>1 rem/yr	<ul style="list-style-type: none"> <li>• Shall wear a leaded apron of 0.25 mm lead equivalent; and</li> </ul>

<b>If the predicted whole body dose rate (effective dose) is:</b>	<b>And the shallow dose rate to the fingers is:</b>	<b>The operator:</b>
		<ul style="list-style-type: none"> <li>• Shall use equipment for special needs patients outside of routine dental office settings only</li> </ul>

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-085, filed 9/7/11, effective 10/8/11.]

**WAC 246-225A-090 X-ray image processing requirements.** Standards in this section are designed to assure that optimal X-ray image quality and diagnostic information are produced so that fewer retakes are needed, and associated patient and operator exposure is minimized.

(1) When performing manual film processing (also known as hand tank processing) registrants or an operator working under the registrant's direction shall:

(a) Use appropriate chemicals for manual film processing as indicated in chemical and film manufacturer's labels and recommendations.

(b) Mix chemicals in accordance with the chemical manufacturer's recommendations.

(c) Periodically add film developer/fixer replenisher based on the recommendations of the chemical or film manufacturer. Solution may be removed from the tank to permit the addition of an adequate volume of replenisher.

(d) Completely replace all manual processing chemicals at least every two months, or follow the manufacturer's recommendations for periodic chemistry replenishment and maintenance, whichever is shorter.

(e) Except when quick developer chemistry is used, post and keep for department inspection, the most recent twelve months of a log that shows when each chemistry change was done and by whom.

(f) Process film to achieve the best image quality by either:

(i) Following the film manufacturer's published temperature and time recommendations for X-ray film development; or

(ii) Developing film according to the temperature-time chart in (g) of this subsection.

(g) For standard developer solution, follow the X-ray film developing time specified for the appropriate developer solution temperature in Table 1 of this section:

Table 1

THERMOMETER READINGS (DEGREES)		MINIMUM DEVELOPING TIMES (MINUTES)
C	F	
27	80	2
	79	2
	78	2 1/2
	77	2 1/2
	76	3
24	75	3
	74	3 1/2
	73	3 1/2
	72	4
22	71	4
	70	4 1/2

THERMOMETER READINGS (DEGREES)		MINIMUM DEVELOPING TIMES (MINUTES)
	69	4 1/2
20	68	5
	67	5 1/2
	66	5 1/2
	65	6
18	64	6 1/2
	63	7
	62	8
	61	8 1/2
16	60	9 1/2

- (h) Use X-ray film developing devices that give:
- (i) The actual temperature of the developer solution;
- (ii) The developing time in minutes and seconds; and
- (iii) An audible or visible signal when developing is complete.

(2) When performing automatic film processing, registrants or an operator working under the registrant's direction shall:

(a) Set up and maintain automatic film processors so that X-ray image density and contrast are optimal;

(b) Follow the film manufacturer's published specifications for time and temperature, and the processor manufacturer's recommendations for type of developer chemistry used. If manufacturer's specifications are not available, the film must be developed using the developer temperatures and immersion times specified in Table 2 of this section:

Table 2

DEVELOPER TEMPERATURE		PROCESSOR DEVELOPER IMMERSION TIME*
°C	°F	Seconds
35	95	20
34.5	94	21
34	93	22
33.5	92	23
33	91	24
32	90	25
31.5	89	26
31	88	27
30.5	87	28
30	86	29
29.5	85	30

\*Immersion time only, no cross-over time included.

(c) Replenish the developer chemistry to create optimal X-ray images by:

(i) Replacing all automatic processor chemicals at least every month, or follow the manufacturer's recommendations for periodic chemistry replenishment and maintenance, whichever is shorter.

(ii) Posting and maintaining a log that shows when each chemistry change was performed and by whom. The most recent twelve months of the log must be kept for department inspection.

(iii) Verifying that the processor delivers an adequate rate of developer replenishment; and

(iv) Verifying that standby replenishment, flood replenishment, or prefixed film processing are done as necessary for facilities with a low X-ray workload.

(3) When developing film, registrants or an operator working under the registrant's direction shall:

(a) Set up darkrooms and daylight film loaders so that film being processed, handled, or stored will be exposed only to light passed through a safelight filter. The filter must be of the type specified by the film manufacturer and must not cause excess fog on X-ray-exposed film. Fog greater than 0.1 optical density is considered unacceptable.

(b) Use daylight loaders in darkened areas or where light is dimmed so that the fog standard in (a) of this subsection is met.

(4) When processing digital images, registrants or an operator working under the registrant's direction shall:

(a) Follow the CR and DR sensor or detector manufacturer's recommendations to achieve adequate diagnostic image quality for the least possible patient exposure.

(b) Process CR phosphor plates using the longest processing time recommended by the manufacturer of the plate processor.

(5) The department may make X-ray film development and darkroom tests as necessary to determine compliance with this section.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-090, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-090, filed 6/26/08, effective 7/27/08.]

**WAC 246-225A-110 Film processing quality assurance.** Registrants found in violation of WAC 246-225A-090 (1)(f) or (2) shall comply with the following quality assurance requirements for a minimum of three months:

(1) Conduct an acceptable quality assurance program that includes weekly tests of film processing to include:

(a) Density of test films; and

(b) Action taken when test film density falls below 15 percent of initial reference levels.

(2) Keep a written or computer log of all periodic quality assurance testing covered in subsection (1) of this section, including the weekly test films.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-110, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-110, filed 6/26/08, effective 7/27/08.]