

**WAC 246-229-0100 What are the recordkeeping requirements for particle accelerator use?**

<b>Records</b>	<b>Retention time</b>
Operator training and qualifications	Two years past last employment/operation
Safety and warning device checks	Two years
Area radiation monitors	Two years (if necessary)
Instrumentation tests	Two years
Smear results	Two years (if necessary)
Qualified expert radiation protection surveys	Life of the accelerator
Electrical circuit diagrams	Life of the accelerator
Permanent log of bypassing interlocks	Life of the accelerator

[Statutory Authority: RCW 70.98.050 and [70.98.]080. WSR 02-14-050, § 246-229-0100, filed 6/27/02, effective 7/28/02.]