

**WAC 246-220-080 Prohibited uses.** (1) Hand-held fluoroscopic screens shall not be used unless listed in the Registry of Sealed Sources and Devices or accepted for certifications by the United States Food and Drug Administration, Center for Devices and Radiological Health.

(2) Shoe-fitting fluoroscopic devices shall not be used.

[Statutory Authority: RCW 70.98.050. WSR 94-01-073, § 246-220-080, filed 12/9/93, effective 1/9/94. Statutory Authority: RCW 43.70.040. WSR 91-02-049 (Order 121), recodified as § 246-220-080, filed 12/27/90, effective 1/31/91; Order 1095, § 402-12-150, filed 2/6/76.]