

WAC 246-817-722 Defibrillator. When anesthetic agents of any kind are administered, the dentist and staff must have access to an AED or defibrillator. The AED or defibrillator must be available and in reach within 60 seconds.

[Statutory Authority: RCW 18.32.0365 and 18.32.640. WSR 24-01-033, § 246-817-722, filed 12/11/23, effective 1/11/24; WSR 10-23-001, § 246-817-722, filed 11/3/10, effective 12/4/10; WSR 09-04-042, § 246-817-722, filed 1/30/09, effective 3/2/09.]