

**WAC 246-225A-080 Special requirements for dental intra-oral radiography.** (1) Registrants using an X-ray system designed for use with an intra-oral image receptor shall use equipment that:

- (a) Limits the SSD to not less than 18 centimeters;
- (b) Limits the X-ray beam so that the beam diameter at the minimum SSD is no greater than 7 centimeters in diameter;
- (c) Has an open-ended position-indicating device; and
- (d) Has shielding included in the beam-limiting device or position-indicating device equivalent to that required for the diagnostic source assembly under WAC 246-225A-060(1).

(2) Registrants shall not use diagnostic dental X-ray systems with a fixed, nominal kVp of less than 55.

[Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 11-19-013, § 246-225A-080, filed 9/7/11, effective 10/8/11; WSR 08-14-074, § 246-225A-080, filed 6/26/08, effective 7/27/08.]