

WAC 246-240-354 Surveys of patients and human research subjects treated with a remote afterloader unit. (1) Before releasing a patient or a human research subject from licensee control, a licensee shall survey the patient or the human research subject and the remote afterloader unit with a portable radiation detection survey instrument to confirm that the source(s) has been removed from the patient or human research subject and returned to the safe shielded position.

(2) A licensee shall retain a record of these surveys in accordance with WAC 246-240-593.

[Statutory Authority: RCW 70.98.050. WSR 06-05-019, § 246-240-354, filed 2/6/06, effective 3/9/06.]