

WAC 296-817-40030 Make sure a record is kept of audiometric tests. You must retain a legible copy of all employee audiograms conducted under this chapter.

Make sure the record includes:

- (1) Name and job classification of the employee;
- (2) Date of the audiogram;
- (3) The examiner's name;
- (4) Date of the last acoustic or exhaustive calibration of the audiometer;
- (5) Employee's most recent noise exposure assessment;
- (6) The background sound pressure levels in audiometric test rooms.

[Statutory Authority: RCW 49.17.010, 49.17.040, 49.17.050. WSR 15-23-086, § 296-817-40030, filed 11/17/15, effective 12/18/15. Statutory Authority: RCW 49.17.010, 49.17.040, 49.17.050, and 49.17.060. WSR 03-11-060, § 296-817-40030, filed 5/19/03, effective 8/1/03.]