

WAC 246-225-100 Radiographic systems—Standby radiation from capacitor energy storage equipment. Radiation emitted from the X-ray tube when the exposure switch or timer is not activated shall not exceed a rate of 2 milliroentgens per hour at 5 centimeters from any accessible surface of the diagnostic source assembly, with the beam-limiting device fully open.

[Statutory Authority: RCW 43.70.040. WSR 91-02-049 (Order 121), recodified as § 246-225-100, filed 12/27/90, effective 1/31/91. Statutory Authority: RCW 70.98.050. WSR 81-01-011 (Order 1570), § 402-28-055, filed 12/8/80; Order 1084, § 402-28-055, filed 1/14/76. Formerly WAC 402-28-050 (part).]