

**WAC 246-240-605 Records of installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.** A licensee shall retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units as required by WAC 246-240-357 for three years. For each installation, maintenance, adjustment and repair, the record must include the date, description of the service, and names of the individuals who performed the work.

[Statutory Authority: RCW 70A.388.040 and 70A.388.110. WSR 22-19-084, § 246-240-605, filed 9/20/22, effective 10/21/22. Statutory Authority: RCW 70.98.050. WSR 06-05-019, § 246-240-605, filed 2/6/06, effective 3/9/06.]