

WAC 448-16-020 Approval of breath test equipment. (1) Pursuant to RCW 46.61.506, the following instruments are approved for the quantitative measurement of alcohol in a person's breath:

- (a) The DataMaster;
- (b) The DataMaster CDM; and
- (c) The Draeger or Dräger Alcotest 9510.

(2) Pursuant to RCW 46.61.506, the following thermometers are approved:

(a) Mercury in glass thermometers with a scale graduated in tenths of a degree measuring a range between 33.5 and 34.5 degrees centigrade.

(b) Digital thermometer system contained within the Guth 2100 wet bath simulator.

[Statutory Authority: RCW 46.61.506. WSR 16-21-025, § 448-16-020, filed 10/10/16, effective 11/10/16; WSR 10-24-066, § 448-16-020, filed 11/30/10, effective 12/31/10; WSR 04-19-144, § 448-16-020, filed 9/22/04, effective 10/23/04.]