

**WAC 246-229-0080 What are the requirements for bypassing safety interlocks?** Bypassing a safety interlock or interlocks is allowed only if:

- (1) Authorized by the radiation safety committee and/or radiation safety officer;
- (2) Recorded in a permanent log and a notice posted at the accelerator control console; and
- (3) The bypass procedure is terminated as soon as possible.

[Statutory Authority: RCW 70.98.050 and [70.98.]080. WSR 02-14-050, § 246-229-0080, filed 6/27/02, effective 7/28/02.]