

WAC 246-240-611 Records of dosimetry equipment used with remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units. (1) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with WAC 246-240-366 for the duration of the license.

(2) For each calibration, intercomparison, or comparison, the record must include:

(a) The date;

(b) The manufacturer's name, model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by WAC 246-240-366;

(c) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and

(d) The names of the individuals who performed the calibration, intercomparison, or comparison.

[Statutory Authority: RCW 70.98.050. WSR 06-05-019, § 246-240-611, filed 2/6/06, effective 3/9/06.]